

SHIFT FILTRATION

TECHNICAL DATA SHEET According to Good Manufacturing Practice (CGMP) standards

V3.0

GRADE: DESCRIPTION: DATE: MNW045047R-SG CELLULOSE NITRATE (ESTER) MEMBRANE FILTERS STERILE. GRIDDED. REEL MARCH 2012

This document is to verify that the designated product has been manufactured in conformance with applicable Current Good Manufacturing Practice (cGMP) standards.

The quality control data given in this document represent the quality of the released lot. These values are the basis for the official release of this material. The Quality Department for quality control of filters has measured the values and assures that they are within the limits that are established in the current specification for this material. The values stated do not represent any internal or external specification for this particular material. This product has passed external-house tests and thus meets Chmlab Group stringent quality control standards. The following is checked on a regular basis:

Membrane Filter Characterization

PACKAGING:	Boxes of	300 units
FACINAOINO.	DOXES OI	Soo units
FORMAT:	Circles Ø	47 mm
I ORMAN	Circles o	7/ 11111

TECHNICAL SPECIFICATIONS:	
Membrane	Cellulose Nitrate (Ester)
Pore size	0.45µm
Membrane Color Grid Color	white black
Package type	Sterile, individually packed. In reel
Wettability	Hydrophilic
Bubble point minimum value, wetted with water	2.5 bar
Thickness	90-140 μm
Sterilization	By autoclaving at 121 °C, γ-radiation 25 KGray or EO
Thermal Stability	
Extractables	With water less than 1%

Other Specifications

Properties	They are ready-to use membranes and save preparatory time. Filter identification and lot number are printed on the box or on each individual enveloped Hydrophilic membrane Very uniform pore structure which ensures homogeneous distribution of particles retained on the filter surface Autoclavable Very high flow rate.	
Applications	Clarification and sterilistation of aqueous solutions Microbiological analysis and particle counting Particle size analysis Pre-filtration ans clarification of samples prior to further analysis Removal of particicles in suspensions to determinate the degree of impurity	
Chemical compatibility	ity See chemical compatibility table on page 114 of our general catalogue http://www.chmlab.com/en/pdf/CHMLABcatalogue.pdf	

www.chmlab.com